

INTERNATIONAL STANDARD



Electronic displays –
Part 1-31: Generic – Practical information on the use of light measuring devices

IEC TR 62977-1-31:2021

<https://standards.iteh.ai/catalog/standards/sist/1385a523-d79f-4db0-8935-cb951a3f6747/iec-tr-62977-1-31-2021>



THIS PUBLICATION IS COPYRIGHT PROTECTED

Copyright © 2022 IEC, Geneva, Switzerland

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either IEC or IEC's member National Committee in the country of the requester. If you have any questions about IEC copyright or have an enquiry about obtaining additional rights to this publication, please contact the address below or your local IEC member National Committee for further information.

IEC Secretariat
3, rue de Varembe
CH-1211 Geneva 20
Switzerland

Tel.: +41 22 919 02 11
info@iec.ch
www.iec.ch

About the IEC

The International Electrotechnical Commission (IEC) is the leading global organization that prepares and publishes International Standards for all electrical, electronic and related technologies.

About IEC publications

The technical content of IEC publications is kept under constant review by the IEC. Please make sure that you have the latest edition, a corrigendum or an amendment might have been published.

IEC publications search - webstore.iec.ch/advsearchform

The advanced search enables to find IEC publications by a variety of criteria (reference number, text, technical committee, ...). It also gives information on projects, replaced and withdrawn publications.

IEC Just Published - webstore.iec.ch/justpublished

Stay up to date on all new IEC publications. Just Published details all new publications released. Available online and once a month by email.

IEC Customer Service Centre - webstore.iec.ch/csc

If you wish to give us your feedback on this publication or need further assistance, please contact the Customer Service Centre: sales@iec.ch.

IEC Products & Services Portal - products.iec.ch

Discover our powerful search engine and read freely all the publications previews. With a subscription you will always have access to up to date content tailored to your needs.

Electropedia - www.electropedia.org

The world's leading online dictionary on electrotechnology, containing more than 22 300 terminological entries in English and French, with equivalent terms in 19 additional languages. Also known as the International Electrotechnical Vocabulary (IEV) online.

IEC TR 62977-1-31:2021

<https://standards.iteh.ai/catalog/standards/sist/1385a523-d79f-4db0-8935-cb951a3f6747/iec-tr-62977-1-31-2021>



INTERNATIONAL STANDARD



Electronic displays –
Part 1-31: Generic – Practical information on the use of light measuring devices

IEC TR 62977-1-31:2021

<https://standards.iteh.ai/catalog/standards/sist/1385a523-d79f-4db0-8935-cb951a3f6747/iec-tr-62977-1-31-2021>

INTERNATIONAL
ELECTROTECHNICAL
COMMISSION

ICS 31.120; 31.260

ISBN 978-2-8322-4572-9

Warning! Make sure that you obtained this publication from an authorized distributor.

REDLINE VERSION



Electronic displays –
Part 1-31: Generic – Practical information on the use of light measuring devices

IEC TR 62977-1-31:2021

<https://standards.iteh.ai/catalog/standards/sist/1385a523-d79f-4db0-8935-cb951a3f6747/iec-tr-62977-1-31-2021>

CONTENTS

FOREWORD	7
INTRODUCTION	9
INTRODUCTION to Amendment 1	9
1 Scope	10
2 Normative references	10
3 Terms, definitions, and abbreviated terms	10
3.1 Terms and definitions	10
3.2 Abbreviated terms	11
4 General information on LMDs for photometry and colorimetry	11
4.1 General	11
4.2 Photometry and colorimetry for electronic displays	11
4.3 LMDs for luminance and chromaticity measurements	12
4.3.1 Configuration of LMDs	12
4.3.2 Input optics of LMDs	15
4.3.3 Electronic system of LMDs	15
4.3.4 Calibration of LMDs	16
4.3.5 Maintenance of LMDs	16
4.4 Setup conditions for measurement	16
4.4.1 LMDs	16
4.4.2 DUTs	16
4.4.3 Environment	16
5 Influence of LMD properties on luminance and chromaticity measurements	16
5.1 General	16
5.2 Repeatability	17
5.2.1 General	17
5.2.2 Example of the repeatability of an LMD	17
5.3 Accuracy	18
5.3.1 General	18
5.3.2 Example of the accuracy of an LMD	18
5.3.3 Linearity	18
5.3.4 Range change	19
5.4 Luminance range	19
5.5 Spectral properties of the spectroradiometer	19
5.5.1 General	19
5.5.2 Wavelength accuracy and spectral bandwidth	19
5.6 Spectral properties of the filter-type luminance meter and colorimeter	26
5.6.1 General	26
5.6.2 Spectral responsivity	27
5.6.3 Methods to reduce the measurement difference	30
5.7 Angular response of LMDs	31
5.7.1 General	31
5.7.2 Subtended angles	31
5.7.3 Consideration of the input optics	32
5.8 Measurement field	33
5.8.1 General	33
5.8.2 Number of pixels within the measurement field	34

5.9	Polarization.....	34
5.9.1	General	34
5.9.2	Polarization dependence of LMDs.....	35
5.10	Temporal synchronization	36
5.10.1	General	36
5.10.2	Temporal synchronization of the LMD and DUT	36
6	Influence of LMD properties on measurements of the optical characteristics of electronic displays	37
6.1	General.....	37
6.2	Contrast ratio	37
6.2.1	General	37
6.2.2	Calculated influence of LMD properties on the contrast ratio measurements	37
6.3	Electro-optical transfer function (EOTF)	40
6.3.1	General	40
6.3.2	Calculated influence of the LMD properties on the EOTF measurements	40
6.4	Chromaticity gamut area	41
6.4.1	General	41
6.4.2	Calculated influence of LMD properties on the chromaticity gamut area measurements	41
6.5	Viewing direction characteristics	43
6.5.1	General	43
6.5.2	Calculated influence of the LMD properties on the viewing direction characteristics measurements	43
6.5.3	Measurement field at an oblique direction.....	45
6.6	Spatial uniformity	46
6.6.1	General	46
6.6.2	Calculated influence of LMD properties on uniformity and non-uniformity measurements	46
6.7	Response time	48
6.7.1	General	48
6.7.2	Measurement of the response time	48
6.8	Flicker.....	51
6.8.1	General	51
6.8.2	Measurement method of the flicker	51
6.8.3	Low-pass filter of LMDs	52
Annex A (informative)	Photometry and colorimetry	54
A.1	General.....	54
A.2	Photometry	54
A.3	Colorimetry	54
A.3.1	General	54
A.3.2	Standard colorimetric observer	55
A.3.3	Tristimulus values.....	55
A.3.4	Chromaticity diagram and colour space	55
Annex B (informative)	Method for reducing the measurement difference of colorimeters	58
B.1	General.....	58
B.2	Matrix calibration methods for colorimeters	58
B.2.1	Matrix calibration process 1: RGB calibration.....	58
B.2.2	Matrix calibration process 2: RGBW calibration	59

Annex C (informative) Input data in Clause 5 and Clause 6, and calculation methods in 5.8 and 6.5.....	61
C.1 General.....	61
C.2 Characteristics of DUTs	61
C.2.1 Spectral radiances of the DUTs	61
C.2.2 Directional characteristic of the DUT.....	62
C.2.3 Temporal modulation characteristics of the DUT	63
C.2.4 EOTF characteristics of the DUTs.....	63
C.2.5 Uniformity characteristics of the DUTs	64
C.3 Characteristics of the filter-type LMDs	65
C.3.1 Spectral responsivities of the filter-type LMDs	65
C.3.2 Specifications of filter-type LMDs.....	67
C.4 Influence of the number of pixels within the measurement field	69
C.5 Validity of the viewing direction dependence obtained by a simplified method	70
Annex D (informative) Instabilities of DUTs in measurement	72
D.1 General.....	72
D.2 DUT instabilities.....	72
Annex E (informative) Influence of spectral stray light and spectral bandwidth on chromaticity measurements of narrow spectral linewidth light	74
E.1 General.....	74
E.2 Example experimental data on spectral stray light.....	74
E.2.1 Measurement method and conditions.....	74
E.2.2 Measurement and calculation data.....	74
E.2.3 Calculation of chromaticity with increasing background level	76
E.3 Influence of spectral bandwidth.....	77
E.3.1 Calculation of chromaticity with increasing spectral bandwidth.....	77
E.3.2 Additional information on spectral properties of spectroradiometers.....	77
Bibliography.....	79
Figure 1 – Block diagrams of three types of LMDs	13
Figure 2 – Example of configurations for the input optics and detector	14
Figure 3 – Example of input optics for the luminance meters.....	15
Figure 4 – Block diagram of a typical electronic system	15
Figure 5 – Examples of the repeatability of an LMD as a function of luminance	17
Figure 6 – Examples of the accuracy of an LMD as a function of luminance	18
Figure 7 – Calculated relative luminance difference as a function of wavelength error	21
Figure 8 – Calculated relative luminance difference as a function of spectral bandwidth	22
Figure 9 – Calculated chromaticity differences as a function of wavelength error	24
Figure 10 – Calculated chromaticity differences as a function of spectral bandwidth	26
Figure 11 – Calculated relative luminance difference as a function of f_1'	28
Figure 12 – Calculated chromaticity differences as a function of f_1' , xyz	30
Figure 13 – Angular aperture and measurement field angle	32
Figure 14 – Calculated relative luminance difference and chromaticity difference as a function of the angular aperture	32
Figure 15 – Diagram of light rays in object space telecentric and non-telecentric optical design.....	33

Figure 16 – Calculated chromaticity difference as a function of the number of pixels	34
Figure 17 – Measured luminance variation as a function of the rotation angle of the polarizer	35
Figure 18 – Calculated relative luminance differences as a function of sampling period	37
Figure 19 – Total measurement times for 0,3 % repeatability (2σ) in three LMDs	41
Figure 20 – Calculated relative difference, ΔGA_{xy} , by spectroradiometers	42
Figure 21 – Calculated relative difference, ΔGA_{xy} , by colorimeters	43
Figure 22 – Cone of light rays for calculating the tristimulus values measured by an LMD with the angular aperture, α , and an optical axis at an inclination angle, θ_{LMD}	44
Figure 23 – Calculated luminance and chromaticity dependence as a function of the inclination angle for the 2°, 6°, and 10° angular apertures.....	45
Figure 24 – Measurement field and test pattern	46
Figure 25 – Calculated non-uniformity difference by the filter-type colorimeters	48
Figure 26 – Measurement setups for response time measurements	49
Figure 27 – Response curves measured at different sampling rates	49
Figure 28 – Measured response subjected to various low-pass filterings.....	50
Figure 29 – Measured response curves switched from the 10 % to 90 % level	51
Figure 30 – Schematic measured temporal luminance modulation of the LCD with a common voltage offset.....	52
Figure 31 – Conceptual pseudo-temporal luminance modulation	53
Figure 32 – Simulated luminance modulations with and without high frequency noise	53
Figure C.1 – Spectral radiances.....	62
Figure C.2 – Inclination angle dependence	63
Figure C.3 – Temporal modulation of the luminance	63
Figure C.4 – EOTF characteristics	64
Figure C.5 – Spectral responsivity of the colorimeter	67
Figure C.6 – Repeatability of three LMD models	68
Figure C.7 – Three positions of a circular measurement field relative to the RGB stripes	69
Figure C.8 – Configuration to measure DUT emission by the LMD at the inclination angle, θ_{LMD}	70
Figure C.9 – Inclination angle dependence calculated by two methods using an LMD with an angular aperture of 10°	71
Figure D.1 – Changes of luminance and chromaticity after switching the grey levels.....	72
Figure D.2 – Changes of luminance after switching from the black to grey level.....	73
Figure E.1 – Measured relative spectral irradiance for each primary input signal.....	75
Figure E.2 – Calculated CIE 1931 chromaticities of simulated monochromatic spectra with various wavelengths with increasing background levels (signal-to-noise ratios)	76
Figure E.3 – Calculated CIE 1931 chromaticities of simulated monochromatic spectra for various wavelengths with increasing spectral bandwidths	78
Table 1 – DUT characteristics for the calculations.....	38
Table 2 – Calculated results of the contrast ratio by three LMDs	39
Table 3 – Calculated durations of the EOTF measurements.....	41
Table 4 – Chromaticity gamut area of three DUTs.....	42

Table 5 – Non-uniformity of DUTs	47
Table 6 – Rise times calculated from a measured response subjected to various low-pass filterings	50
Table 7 – Rise times measured by LMDs of various $V(\lambda)$ fidelities	51
Table C.1 – Measured luminance and chromaticity at the nine positions of DUT-1	64
Table C.2 – Measured luminance and chromaticity at the nine positions of DUT-2	65
Table C.3 – Specifications of filter-type LMDs.....	68
Table D.1 – Luminance and chromaticity transition	72
Table D.2 – Luminance changes in two measurements	73
Table E.1 – Calculated CIE 1931 chromaticity differences between measurements without and with spectral stray light correction	75

iTeh STANDARD PREVIEW
(standards.iteh.ai)

IEC TR 62977-1-31:2021

<https://standards.iteh.ai/catalog/standards/sist/1385a523-d79f-4db0-8935-cb951a3f6747/iec-tr-62977-1-31-2021>

INTERNATIONAL ELECTROTECHNICAL COMMISSION

ELECTRONIC DISPLAYS –

Part 1-31: Generic – Practical information on the use of light measuring devices

FOREWORD

- 1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as "IEC Publication(s)"). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
- 3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
- 4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.
- 5) IEC itself does not provide any attestation of conformity. Independent certification bodies provide conformity assessment services and, in some areas, access to IEC marks of conformity. IEC is not responsible for any services carried out by independent certification bodies.
- 6) All users should ensure that they have the latest edition of this publication.
- 7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
- 8) Attention is drawn to the Normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
- 9) Attention is drawn to the possibility that some of the elements of this IEC Publication may be the subject of patent rights. IEC shall not be held responsible for identifying any or all such patent rights.

This consolidated version of the official IEC Standard and its amendment has been prepared for user convenience.

IEC TR 62977-1-31 edition 1.1 contains the first edition (2021-04) [documents 110/1258/DTR and 110/1281A/RVDTR] and its amendment 1 (2022-03) [documents 110/1380/DTR and 110/1404A/DVDTR].

In this Redline version, a vertical line in the margin shows where the technical content is modified by amendment 1. Additions are in green text, deletions are in strikethrough red text. A separate Final version with all changes accepted is available in this publication.

IEC TR 62977-1-31 has been prepared by IEC technical committee 110: Electronic displays. It is a Technical Report.

This document was drafted in accordance with ISO/IEC Directives, Part 2, and developed in accordance with ISO/IEC Directives, Part 1 and ISO/IEC Directives, IEC Supplement, available at www.iec.ch/members_experts/refdocs. The main document types developed by IEC are described in greater detail at www.iec.ch/standardsdev/publications.

A list of all parts in the IEC 62977 series, published under the general title *Electronic displays*, can be found on the IEC website.

The committee has decided that the contents of the base publication and its amendment will remain unchanged until the stability date indicated on the IEC web site under webstore.iec.ch in the data related to the specific publication. At this date, the publication will be

- reconfirmed,
- withdrawn,
- replaced by a revised edition, or
- amended.

IMPORTANT – The 'colour inside' logo on the cover page of this publication indicates that it contains colours which are considered to be useful for the correct understanding of its contents. Users should therefore print this document using a colour printer.

(standards.iteh.ai)

IEC TR 62977-1-31:2021

<https://standards.iteh.ai/catalog/standards/sist/1385a523-d79f-4db0-8935-cb951a3f6747/iec-tr-62977-1-31-2021>

INTRODUCTION

Measurements of the optical characteristics of electronic displays are primarily affected by three factors: measuring procedures, displays (devices under test: DUTs), and light measuring devices (LMDs), for which there are many international standards supporting consistent and comparable measurements. Most of them, however, provide only limited information on LMDs, making it difficult to appropriately select and use the LMD for the measurement objective. The purpose of this document is to provide best practices and suggestions which are missing in the standards.

This document addresses how the major properties of a typical LMD affect the measurement results. It is often impractical and unnecessary to consider the influences of all properties of LMDs and all characteristics of DUTs as well as their interactions and influences on the measurement results. Therefore, the multiple interaction effects that exist are beyond the scope of this document. Due to the rapid innovation and abundance of LMDs, covering all types of LMDs is also outside the objectives of this document.

INTRODUCTION to Amendment 1

This document provides additional information to IEC TR 62977-1-31:2021 regarding the influence of spectral stray light and spectral bandwidth of a spectroradiometer on chromaticity measurements. It is described in Annex E.

This document also provides the corrections of editorial errors of IEC TR 62977-1-31:2021. The corrections are:

- Typos are fixed:
 - “fiber” and “ $x(\lambda)$, $y(\lambda)$, $z(\lambda)$ ” is replaced with “fibre” and “ $\bar{x}(\lambda)$, $\bar{y}(\lambda)$, $\bar{z}(\lambda)$ ”, respectively in Figure 2,
 - “(%)” in the label of vertical axis is removed in Figure 9, Figure 10, and Figure 12,
 - “0” label of the tick mark of vertical axis is replaced with “1” in Figure C.4.
- The lists for Formula (A.1) and Formula (B.1) are aligned.

ELECTRONIC DISPLAYS –

Part 1-31: Generic – Practical information on the use of light measuring devices

1 Scope

This part of IEC 62977 provides practical information on light measuring devices (luminance meters, colorimeters, and spectroradiometers) with luminance measuring optics for the characterization of electronic displays.

2 Normative references

There are no normative references in this document.

3 Terms, definitions, and abbreviated terms

3.1 Terms and definitions

For the purposes of this document, the following terms and definitions apply.

ISO and IEC maintain terminological databases for use in standardization at the following addresses:

- ISO Online browsing platform: available at <http://www.iso.org/obp>
- IEC Electropedia: available at <http://www.electropedia.org/>

NOTE CIE Electronic international lighting vocabulary (e-ILV) is also available at <http://cie.co.at/e-ilv>.

3.1.1

repeatability

<of an LMD> closeness of agreement between indications or measured quantity values obtained by replicated measurements over a short period of time using a specific LMD under conditions specified by the LMD manufacturer

Note 1 to entry: Repeatability of an LMD is usually expressed numerically by statistical quantities, such as standard deviation, variance, or coefficient of variation (relative standard deviation) under the specified conditions of measurement.

Note 2 to entry: The influence on measurement repeatability caused by fluctuations of the measured light source and by the measurement procedure is assumed to be negligible when the manufacturer specifies the repeatability of an LMD. Manufacturers often specify the type of light source and measurement conditions used for determining the repeatability of an LMD.

Note 3 to entry: Measurement precision is the closeness of agreement between indications or measured quantity values obtained by replicate measurements on the same or similar objects under specified conditions. Measurement repeatability is measurement precision under a set of repeatability conditions of measurement that includes the same measurement procedures, same operators, same measuring system, same operating conditions, same location, and replicate measurements on the same or similar objects over a short period of time. Measurement reproducibility is measurement precision under a set of reproducibility conditions of measurement that includes different locations, operators, measuring systems, and replicate measurements on the same or similar objects [1], [2]¹.

¹ Numbers in square brackets refer to the Bibliography.

3.1.2

accuracy

<of an LMD> difference between a measured quantity value and an accepted reference value when using a specific LMD under conditions specified by the LMD manufacturer

Note 1 to entry: This term is a quantity with a numerical value and is usually expressed as a range specification.

Note 2 to entry: The accepted reference value is a value that serves as an agreed-upon reference for comparison, and which is derived as:

- a) a theoretical or established value, based on scientific principles;
- b) an assigned or certified value, based on experimental work of some national or international organization;
- c) a consensus or certified value, based on collaborative experimental work under the auspices of a scientific or engineering group;
- d) (when a), b) and c) are not available) the expectation of the (measurable) quantity, i.e. the mean of a specified population of measurements [3].

Note 3 to entry: The influence on measurement accuracy caused by fluctuations of the measured light source and by the measurement procedure is assumed to be negligible when the manufacturer specifies the accuracy of an LMD. Manufacturers often specify the type of light source and other measurement conditions used for determining the accuracy of an LMD.

Note 4 to entry: Measurement accuracy is the closeness of agreement between a measured quantity value and the true quantity value of a measurand [1], [2]. The accuracy of measurement is not a quantity value while the accuracy of an LMD is a quantity value; thus, the term "accuracy" conventionally used for the specification of LMDs means something different than that used for measurement.

3.2 Abbreviated terms

CIE	Commission Internationale de l'Éclairage (International Commission on Illumination)
CMF	colour-matching function
DUT	device under test
EOTF	electro-optical transfer function
LCD	liquid crystal display
LED	light emitting diode
LMD	light measuring device
ND	neutral density
OLED	organic light emitting diode
PWM	pulse width modulation
RGB	red, green, and blue
RGBW	red, green, blue, and white
Vsync	vertical synchronizing signal

4 General information on LMDs for photometry and colorimetry

4.1 General

Clause 4 describes the principles of photometry and colorimetry, configuration, calibration, and maintenance of LMDs, as well as setup conditions for measurement.

4.2 Photometry and colorimetry for electronic displays

Photometry is the measurement of quantities referring to radiation as evaluated according to a given spectral luminous efficiency (see IEV 845-25-013). Colorimetry is the measurement of colour stimuli based on a set of conventions (see IEV 845-25-014). Details on the calculation formulae and specific conditions applied to electronic display measurement are shown in Annex A.