

SLOVENSKI STANDARD

SIST EN 61094-1:2001

01-september-2001

Measurement microphones - Part 1: Specifications for laboratory standard microphones (IEC 61094-1:2000)

Measurement microphones -- Part 1: Specifications for laboratory standard microphones

Messmikrofone -- Teil 1: Anforderungen an Laboratoriums-Normalmikrofone

Microphones de mesure -- Partie 1: Spécifications des microphones étalons de laboratoire

iTeh STANDARD PREVIEW
(standards.iteh.ai)

Ta slovenski standard je istoveten z: **EN 61094-1:2000**
<https://standards.iteh.ai/catalog/standards/sist/cca6953a-1bce-4fd8-b3b6-221fefac1b04/sist-en-61094-1-2001>

ICS:

17.140.50	Elektroakustika	Electroacoustics
33.160.50	Pribor	Accessories

SIST EN 61094-1:2001

en

iTeh STANDARD PREVIEW
(standards.iteh.ai)

SIST EN 61094-1:2001

<https://standards.iteh.ai/catalog/standards/sist/cea6953a-1be6-4fd8-b3b6-221fefac1b04/sist-en-61094-1-2001>

EUROPEAN STANDARD

EN 61094-1

NORME EUROPÉENNE

EUROPÄISCHE NORM

November 2000

ICS 17.140.50

Supersedes EN 61094-1:1994

English version

Measurement microphones
Part 1: Specifications for laboratory standard microphones
(IEC 61094-1:2000)

Microphones de mesure
Partie 1: Spécifications des microphones
étalons de laboratoire
(CEI 61094-1:2000)

Messmikrofone
Teil 1: Anforderungen an Laboratoriums-
Normalmikrofone
(IEC 61094-1:2000)

iTeh STANDARD PREVIEW
(standards.iteh.ai)

This European Standard was approved by CENELEC on 2000-09-01. CENELEC members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration.

Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the Central Secretariat or to any CENELEC member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CENELEC member into its own language and notified to the Central Secretariat has the same status as the official versions.

CENELEC members are the national electrotechnical committees of Austria, Belgium, Czech Republic, Denmark, Finland, France, Germany, Greece, Iceland, Ireland, Italy, Luxembourg, Netherlands, Norway, Portugal, Spain, Sweden, Switzerland and United Kingdom.

CENELEC

European Committee for Electrotechnical Standardization
Comité Européen de Normalisation Electrotechnique
Europäisches Komitee für Elektrotechnische Normung

Central Secretariat: rue de Stassart 35, B - 1050 Brussels

Foreword

The text of document 29/452/FDIS, future edition 2 of IEC 61094-1, prepared by IEC TC 29, Electroacoustics, was submitted to the IEC-CENELEC parallel vote and was approved by CENELEC as EN 61094-1 on 2000-09-01.

This European Standard supersedes EN 61094-1:1994.

The following dates were fixed:

- latest date by which the EN has to be implemented
at national level by publication of an identical
national standard or by endorsement (dop) 2001-06-01
- latest date by which the national standards conflicting
with the EN have to be withdrawn (dow) 2003-09-01

Endorsement notice

The text of the International Standard IEC 61094-1:2000 was approved by CENELEC as a European Standard without any modification.

iTeh STANDARD PREVIEW
(standards.iteh.ai)

SIST EN 61094-1:2001

<https://standards.iteh.ai/catalog/standards/sist/cea6953a-1be6-4fd8-b3b6-221fefac1b04/sist-en-61094-1-2001>

Annex ZA
(normative)**Normative references to international publications
with their corresponding European publications**

This European Standard incorporates by dated or undated reference, provisions from other publications. These normative references are cited at the appropriate places in the text and the publications are listed hereafter. For dated references, subsequent amendments to or revisions of any of these publications apply to this European Standard only when incorporated in it by amendment or revision. For undated references the latest edition of the publication referred to applies (including amendments).

NOTE When an international publication has been modified by common modifications, indicated by (mod), the relevant EN/HD applies.

<u>Publication</u>	<u>Year</u>	<u>Title</u>	<u>EN/HD</u>	<u>Year</u>
IEC 60050-801	1994	International Electrotechnical Vocabulary (IEV) - Chapter 801: Acoustics and electroacoustics	-	-
ASME B1.1	1989	Unified inch screw threads (UN and UNR thread form)	-	-

iTeh STANDARD PREVIEW
(standards.iteh.ai)

SIST EN 61094-1:2001

<https://standards.iteh.ai/catalog/standards/sist/cea6953a-1be6-4fd8-b3b6-221fefac1b04/sist-en-61094-1-2001>

iTeh STANDARD PREVIEW
(standards.iteh.ai)

SIST EN 61094-1:2001

<https://standards.iteh.ai/catalog/standards/sist/cea6953a-1be6-4fd8-b3b6-221fefac1b04/sist-en-61094-1-2001>

NORME INTERNATIONALE INTERNATIONAL STANDARD

**CEI
IEC**

61094-1

Deuxième édition
Second edition
2000-07

Microphones de mesure –

Partie 1: Spécifications des microphones étalons de laboratoire

iTeh STANDARD PREVIEW

Measurement microphones –

Part 1: [SIST EN 61094-1:2001](https://standards-iteh.ai/catalog/standards/sist/6a6953a-1be6-4f38-b3b6-2211e1ac1604/sist-en-61094-1-2001) Specifications for laboratory standard microphones

© IEC 2000 Droits de reproduction réservés — Copyright - all rights reserved

Aucune partie de cette publication ne peut être reproduite ni utilisée sous quelque forme que ce soit et par aucun procédé, électronique ou mécanique, y compris la photocopie et les microfilms, sans l'accord écrit de l'éditeur.

No part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from the publisher.

International Electrotechnical Commission
Telefax: +41 22 919 0300

3, rue de Varembe Geneva, Switzerland
e-mail: inmail@iec.ch IEC web site <http://www.iec.ch>



Commission Electrotechnique Internationale
International Electrotechnical Commission
Международная Электротехническая Комиссия

CODE PRIX
PRICE CODE

N

Pour prix, voir catalogue en vigueur
For price, see current catalogue

CONTENTS

	Page
FOREWORD	5
Clause	
1 Scope	7
2 Normative references.....	7
3 Terms and definitions.....	7
4 Reference environmental conditions	13
5 Classification of laboratory standard microphone	13
5.1 General.....	13
5.2 Type designation	15
6 Characteristics of laboratory standard microphones	15
6.1 Sensitivity	15
6.2 Acoustic impedance.....	15
6.2.1 General.....	15
6.2.2 Equivalent volume of a microphone.....	15
6.3 Upper limit of the dynamic range of a microphone.....	17
6.4 Static pressure dependence of microphone sensitivity	17
6.5 Temperature dependence of microphone sensitivity.....	17
6.6 Humidity dependence of microphone sensitivity	17
6.7 Electrical insulation resistance.....	19
6.8 Stability of microphone sensitivity	19
6.9 Pressure-equalizing leakage	19
7 Specifications	19
7.1 Mechanical dimensions.....	19
7.2 Ground shield reference configuration	23
7.3 Electroacoustical specifications	25
7.4 Identification markings	27

INTERNATIONAL ELECTROTECHNICAL COMMISSION

MEASUREMENT MICROPHONES –

Part 1: Specifications for laboratory standard microphones

FOREWORD

- 1) The IEC (International Electrotechnical Commission) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of the IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, the IEC publishes International Standards. Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. The IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of the IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested National Committees.
- 3) The documents produced have the form of recommendations for international use and are published in the form of standards, technical specifications, technical reports or guides and they are accepted by the National Committees in that sense.
- 4) In order to promote international unification, IEC National Committees undertake to apply IEC International Standards transparently to the maximum extent possible in their national and regional standards. Any divergence between the IEC Standard and the corresponding national or regional standard shall be clearly indicated in the latter.
- 5) The IEC provides no marking procedure to indicate its approval and cannot be rendered responsible for any equipment declared to be in conformity with one of its standards.
- 6) Attention is drawn to the possibility that some of the elements of this International Standard may be the subject of patent rights. The IEC shall not be held responsible for identifying any or all such patent rights.

International Standard IEC 61094-1 has been prepared by IEC technical committee 29: Electroacoustics.

This second edition cancels and replaces the first edition and corrigendum published in 1992. This second edition constitutes a technical revision.

The text of this standard is based on the first edition, the corrigendum and the following documents:

FDIS	Report on voting
29/452/FDIS	29/461/RVD

Full information on the voting for the approval of this standard can be found in the report on voting indicated in the above table.

This publication has been drafted in accordance with the ISO/IEC Directives, Part 3.

The committee has decided that this publication remains valid until 2005. At this date, in accordance with the committee's decision, the publication will be:

- reconfirmed;
- withdrawn;
- replaced by a revised edition, or
- amended.

MEASUREMENT MICROPHONES –

Part 1: Specifications for laboratory standard microphones

1 Scope

This part of IEC 61094 specifies mechanical dimensions and certain electroacoustic characteristics for condenser microphones used as laboratory standards for the realization of the unit of sound pressure and for sound pressure measurements of the highest attainable accuracy. The specifications are intended to ensure that primary calibration by the reciprocity method can be readily carried out.

This part also establishes a system for classifying laboratory standard condenser microphones into a number of types according to their dimensions and properties in order to facilitate the specification of calibration methods, the conducting of inter-laboratory comparisons involving the calibration of the same microphones in different laboratories, and the interchangeability of microphones in a given calibration system.

2 Normative references

The following normative documents contain provisions which, through reference in this text, constitute provisions of this part of IEC 61094. For dated references, subsequent amendments to, or revisions of, any of these publications do not apply. However parties to agreements based on this part of IEC 61094 are encouraged to investigate the possibility of applying the most recent editions of the normative documents indicated below. For undated references, the latest edition of the normative document referred to applies. Members of ISO and IEC maintain registers of currently valid International Standards.

IEC 60050(801):1994, *International Electrotechnical Vocabulary (IEV) – Chapter 801: Acoustics and electroacoustics*

ASME B1.1:1989, *Unified inch screw threads (UN and UNR thread form)* ¹

3 Terms and definitions

For the purposes of this part of IEC 61094, the following definitions apply.

Remark – The underlined symbols are complex quantities.

3.1

condenser microphones

microphone that operates by variation of electrical capacitance

[IEV 801-26-13].

NOTE Only condenser microphones operating by a virtually constant charge obtained from an external polarizing voltage applied from a source of suitably high internal resistance are considered.

¹ (American Society of Mechanical Engineers) Reference is given to ASME B1.1 in the absence of an equivalent international standard.

3.2**laboratory standard microphone**

condenser microphone capable of being calibrated to a very high accuracy by a primary method such as the closed coupler reciprocity method, and meeting certain severe requirements on mechanical dimensions and electroacoustical characteristics, especially with respect to stability in time and dependence on environmental conditions

3.3**open-circuit voltage**

alternating voltage appearing at the electrical output terminals of a microphone as measured by the insert voltage technique when the microphone is attached to the ground shield configuration specified in 7.2 but is otherwise unloaded

Unit: volt, V

NOTE Owing to the capacitive nature of the microphone, the voltage at the electrical terminals depends on the electrical load presented by the mechanical and electrical attachment of the microphone to a preamplifier. For this reason, preamplifiers used for measuring the open-circuit voltage of a microphone should fulfill the requirements of 7.2.

3.4**pressure sensitivity of a microphone**

for a sinusoidal signal of given frequency and for given environmental conditions, the quotient of the open-circuit voltage of the microphone by the sound pressure acting over the exposed surface of the diaphragm (i.e. at the acoustical terminals of the microphone), the sound pressure being uniformly applied over the surface of the diaphragm. This quotient is a complex quantity, but when phase information is of no interest the pressure sensitivity may denote its modulus only

Unit: volt per pascal, V/Pa

(standards.iteh.ai)

3.5**pressure sensitivity level of a microphone**

logarithm of the ratio of the modulus of the pressure sensitivity $|M_p|$ to a reference sensitivity. The pressure sensitivity level in decibels is $20 \lg (|M_p| / M_r)$, where the reference sensitivity M_r is 1 V/Pa

Unit: decibel, dB

3.6**free-field sensitivity of a microphone**

for a sinusoidal plane progressive sound wave of given frequency, for a specified direction of incidence, and for given environmental conditions, the quotient of the open-circuit voltage of the microphone by the sound pressure that would exist at the position of the acoustic centre of the microphone in the absence of the microphone. This quotient is a complex quantity, but when phase information is of no interest, the free-field sensitivity may denote its modulus only

Unit: volt per pascal, V/Pa

NOTE 1 At frequencies sufficiently low for the disturbance of the sound field by the microphone to be negligible, the free-field sensitivity approaches the pressure sensitivity (see 6.9 for practical limitations).

NOTE 2 The position of the acoustic centre is a function of frequency.

3.7**free-field sensitivity level of a microphone**

logarithm of the ratio of the modulus of the free-field sensitivity $|M_f|$ to a reference sensitivity. The free-field sensitivity level in decibels is $20 \lg (|M_f| / M_r)$, where the reference sensitivity M_r is 1 V/Pa

Unit: decibel, dB