

---

**Medicinska električna oprema - 2-71. del: Posebne zahteve za osnovno varnost in bistvene lastnosti funkcionalne opreme spektrometra v bližnjem infrardečem spektru (IEC 80601-2-71:2025)**

Medical electrical equipment - Part 2-71: Particular requirements for the basic safety and essential performance of functional near-infrared spectroscopy (NIRS) equipment (IEC 80601-2-71:2025)

Medizinische elektrische Geräte - Teil 2-71: Besondere Festlegungen für die Sicherheit einschließlich der wesentlichen Leistungsmerkmale von funktionalen Oximetriegeräten (IEC 80601-2-71:2025)

Appareils électromédicaux - Partie 2-71: Exigences particulières pour la sécurité de base et les performances essentielles des appareils de spectroscopie dans le proche infrarouge (NIRS) fonctionnelle (IEC 80601-2-71:2025)

**Ta slovenski standard je istoveten z: EN IEC 80601-2-71:2025**

---

**ICS:**

11.040.55      Diagnostična oprema      Diagnostic equipment

**SIST EN IEC 80601-2-71:2025**      en



EUROPEAN STANDARD  
NORME EUROPÉENNE  
EUROPÄISCHE NORM

**EN IEC 80601-2-71**

February 2025

ICS 11.040.55

Supersedes EN IEC 80601-2-71:2018

English Version

**Medical electrical equipment - Part 2-71: Particular requirements  
for the basic safety and essential performance of functional near-  
infrared spectroscopy (NIRS) equipment  
(IEC 80601-2-71:2025)**

Appareils électromédicaux - Partie 2-71: Exigences  
particulières pour la sécurité de base et les performances  
essentielles des appareils de spectroscopie dans le proche  
infrarouge (NIRS) fonctionnelle  
(IEC 80601-2-71:2025)

Medizinische elektrische Geräte - Teil 2-71: Besondere  
Festlegungen für die Sicherheit einschließlich der  
wesentlichen Leistungsmerkmale von funktionalen  
Oximetriegeräten  
(IEC 80601-2-71:2025)

This European Standard was approved by CENELEC on 2025-02-18. CENELEC members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration.

Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CENELEC member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CENELEC member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CENELEC members are the national electrotechnical committees of Austria, Belgium, Bulgaria, Croatia, Cyprus, the Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, the Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Türkiye and the United Kingdom.



European Committee for Electrotechnical Standardization  
Comité Européen de Normalisation Electrotechnique  
Europäisches Komitee für Elektrotechnische Normung

**CEN-CENELEC Management Centre: Rue de la Science 23, B-1040 Brussels**

**EN IEC 80601-2-71:2025 (E)****European foreword**

The text of document 62D/2169/FDIS, future edition 2 of IEC 80601-2-71, prepared by SC 62D "Particular medical equipment, software, and systems" of IEC/TC 62 "Medical equipment, software, and systems" was submitted to the IEC-CENELEC parallel vote and approved by CENELEC as EN IEC 80601-2-71:2025.

The following dates are fixed:

- latest date by which the document has to be implemented at national level by publication of an identical national standard or by endorsement (dop) 2026-02-28
- latest date by which the national standards conflicting with the document have to be withdrawn (dow) 2028-02-29

This document supersedes EN IEC 80601-2-71:2018 and all of its amendments and corrigenda (if any).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CENELEC shall not be held responsible for identifying any or all such patent rights.

Any feedback and questions on this document should be directed to the users' national committee. A complete listing of these bodies can be found on the CENELEC website.

**Endorsement notice**  
(<https://standards.iteh.ai>)  
**Document Preview**

The text of the International Standard IEC 80601-2-71:2025 was approved by CENELEC as a European Standard without any modification.

In the official version, for Bibliography, the following notes have to be added for the standard indicated:

ISO 80601-2-85:2021	NOTE	Approved as EN ISO 80601-2-85:2021 (not modified)
ISO 80601-2-61:2017	NOTE	Approved as EN ISO 80601-2-61:2019 (not modified)
IEC 60601-1-10	NOTE	Approved as EN 60601-1-10
IEC 60601-2-57:2023	NOTE	Approved as EN IEC 60601-2-57:— <sup>1</sup> (not modified)
ISO 14159:2002	NOTE	Approved as EN ISO 14159:2008 (not modified)

<sup>1</sup> Under preparation. Stage at the time of publication: FprEN IEC 60601-2-57:2023.

## Annex ZA (normative)

### Normative references to international publications with their corresponding European publications

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

NOTE 1 Where an International Publication has been modified by common modifications, indicated by (mod), the relevant EN/HD applies.

NOTE 2 Up-to-date information on the latest versions of the European Standards listed in this annex is available here: [www.cencenelec.eu](http://www.cencenelec.eu).

*Annex ZA of EN 60601-1:2006<sup>2</sup>, applies, except as follows:*

*Add:*

<u>Publication</u>	<u>Year</u>	<u>Title</u>	<u>EN/HD</u>	<u>Year</u>
IEC 60601-1	2005	Medical electrical equipment - Part 1: General requirements for basic safety and essential performance	EN 60601-1	2006
-	-	( <a href="https://standards.iteh.ai">https://standards.iteh.ai</a> )	+ AC	2010
+ A1	2012		+ A1	2013
-	-		+ AC	2014
-	-		+ A12	2014
+ A2	2020		+ A2	2021
-	-		+ AC	2022
-	-		+ A13	2024
IEC 60825-1	2014	Safety of laser products - Part 1: Equipment classification and requirements	EN 60825-1	2014
-	-		+ A11	2021
-	-		+ AC	2017-06
IEC 62471 (mod)	2006	Photobiological safety of lamps and lamp systems	EN 62471	2008
IEC 62570	2014	Standard practice for marking medical devices and other items for safety in the magnetic resonance environment	EN 62570	2015

<sup>2</sup> As impacted by EN 60601-1:2006/AC:2010, EN 60601-1:2006/A1:2013, EN 60601-1:2006/A1:2013/AC:2014, EN 60601-1:2006/A12:2014, EN 60601-1:2006/A2:2021, EN 60601-1:2006/AC:2022-12 and EN 60601-1:2006/A13:2024.

**EN IEC 80601-2-71:2025 (E)**

<u>Publication</u>	<u>Year</u>	<u>Title</u>	<u>EN/HD</u>	<u>Year</u>
ISO 17664-1	2021	Processing of health care products - Information to be provided by the medical device manufacturer for the processing of medical devices - Part 1: Critical and semi-critical medical devices	EN ISO 17664-1	2021
ISO 17664-2	2021	Processing of health care products - Information to be provided by the medical device manufacturer for the processing of medical devices - Part 2: Non-critical medical devices	EN ISO 17664-2	2023
ISO 20417	2021	Medical devices - Information to be supplied by the manufacturer	EN ISO 20417	2021

iTeh Standards  
(<https://standards.iteh.ai>)  
Document Preview

[SIST EN IEC 80601-2-71:2025](https://standards.iteh.ai/catalog/standards/sist/6e3ca6cb-172e-442e-8676-a698673c16ba/sist-en-iec-80601-2-71-2025)

<https://standards.iteh.ai/catalog/standards/sist/6e3ca6cb-172e-442e-8676-a698673c16ba/sist-en-iec-80601-2-71-2025>



IEC 80601-2-71

Edition 2.0 2025-01

# INTERNATIONAL STANDARD

## NORME INTERNATIONALE

**Medical electrical equipment –**  
**Part 2-71: Particular requirements for the basic safety and essential performance**  
**of functional near-infrared spectroscopy (NIRS) equipment**

**Appareils électromédicaux –**  
**Partie 2-71: Exigences particulières pour la sécurité de base et les performances**  
**essentiels des appareils de spectroscopie dans le proche infrarouge (NIRS)**  
**fonctionnelle**

INTERNATIONAL  
ELECTROTECHNICAL  
COMMISSION

COMMISSION  
ELECTROTECHNIQUE  
INTERNATIONALE

ICS 11.040.55

ISBN 978-2-8327-0096-9


**Warning! Make sure that you obtained this publication from an authorized distributor.**  
**Attention! Veuillez vous assurer que vous avez obtenu cette publication via un distributeur agréé.**

## CONTENTS

FOREWORD.....	4
INTRODUCTION.....	7
201.1 Scope, object and related standards.....	8
201.2 Normative references .....	10
201.3 Terms and definitions .....	10
201.4 General requirements.....	14
201.5 General requirements for testing ME EQUIPMENT .....	14
201.6 Classification of ME EQUIPMENT and ME SYSTEMS.....	14
201.7 ME EQUIPMENT identification, MARKING and documents .....	14
201.8 Protection against electrical HAZARDS from ME EQUIPMENT .....	16
201.9 Protection against MECHANICAL HAZARDS of ME EQUIPMENT and ME SYSTEMS.....	16
201.10 Protection against unwanted and excessive radiation HAZARDS .....	16
201.11 Protection against excessive temperatures and other HAZARDS .....	16
201.12 Accuracy of controls and instruments and protection against hazardous outputs .....	18
201.13 HAZARDOUS SITUATIONS and fault conditions for ME EQUIPMENT.....	29
201.14 PROGRAMMABLE ELECTRICAL MEDICAL SYSTEMS (PEMS) .....	29
201.15 Construction of ME EQUIPMENT .....	29
201.16 ME SYSTEMS .....	30
201.17 Electromagnetic compatibility of ME EQUIPMENT and ME SYSTEMS .....	30
201.101 Requirements for the FUNCTIONAL NIRS EQUIPMENT ACCESSORIES .....	30
202 Electromagnetic disturbances – Requirements and tests.....	30
206 Usability.....	31
Annexes .....	32
Annex C (informative) Guide to MARKING and labelling requirements for ME EQUIPMENT and ME SYSTEMS .....	33
Annex AA (informative) Particular guidance and rationale .....	35
ANNEX BB (normative) Evaluating ME EQUIPMENT performance using the FUNCTIONAL NIRS PHANTOM.....	39
ANNEX CC (informative) Skin temperature at the FUNCTIONAL NIRS PROBE.....	48
Annex DD (informative) Reference to the IMDRF essential principles and labelling guidance .....	49
Bibliography.....	51
Index of defined terms .....	55
Figure 201.101 – FULL WIDTH AT HALF MAXIMUM OF SPECTRAL POWER DISTRIBUTION .....	11
Figure 201.102 – Measurement of the AVERAGE OPTICAL POWER.....	19
Figure 201.103 – Measurement of PEAK WAVELENGTH and FWHM .....	20
Figure 201.104 – Measurement of the signal stability.....	22
Figure 201.105 – Measurement of the RESPONSE TIME .....	23
Figure 201.106 – Rise time and fall time in the RESPONSE TIME.....	24
Figure 201.107 – Measurement of the signal-to-noise ratio of the detected light intensity .....	25



Figure 201.108 – Measurement of signal-to-noise ratio of the PATHLENGTH-DEPENDENT HAEMOGLOBIN CHANGE .....	27
Figure 201.109 – Measurement of SIGNAL CROSS-TALK .....	28
Figure BB.1 – The FUNCTIONAL NIRS PHANTOM in two states with different detected light intensities .....	42
Figure BB.2 – FUNCTIONAL NIRS PHANTOM measurement using the reference system .....	43
Figure BB.3 – FUNCTIONAL NIRS PHANTOM measurement using the ME EQUIPMENT to be evaluated .....	43
Figure BB.4 – Schematic for measurement of OPTICAL LOSS .....	47
Table 201.101 – Distributed ESSENTIAL PERFORMANCE requirements .....	14
Table 201.102 – Performance tests employing the FUNCTIONAL NIRS EQUIPMENT or attenuator and the required OPTICAL LOSS .....	19
Table 201.C.101 – MARKING on the outside of FUNCTIONAL NIRS EQUIPMENT or their parts .....	33
Table 201.C.102 – ACCOMPANYING DOCUMENTS general .....	33
Table 201.C.103 – INSTRUCTIONS FOR USE .....	34
Table 201.C.104 – TECHNICAL DESCRIPTION .....	34
Table DD.1 – Correspondence between this document and the IMDRF essential principles .....	49
Table DD.2 – Correspondence between this document and the IMDRF labelling principles .....	50

  
 (https://standards.iteh.ai)  
 Document Preview

SIST EN IEC 80601-2-71:2025

<https://standards.iteh.ai/catalog/standards/sist/6e3ca6cb-172e-442e-8676-a698673c16ba/sist-en-iec-80601-2-71-2025>

## INTERNATIONAL ELECTROTECHNICAL COMMISSION

**MEDICAL ELECTRICAL EQUIPMENT –****Part 2-71: Particular requirements for  
the basic safety and essential performance of  
functional near-infrared spectroscopy (NIRS) equipment****FOREWORD**

- 1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as "IEC Publication(s)"). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
- 3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
- 4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.
- 5) IEC itself does not provide any attestation of conformity. Independent certification bodies provide conformity assessment services and, in some areas, access to IEC marks of conformity. IEC is not responsible for any services carried out by independent certification bodies.
- 6) All users should ensure that they have the latest edition of this publication.
- 7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
- 8) Attention is drawn to the Normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
- 9) IEC and ISO draw attention to the possibility that the implementation of this document may involve the use of (a) patent(s). IEC and ISO take no position concerning the evidence, validity or applicability of any claimed patent rights in respect thereof. As of the date of publication of this document, IEC and ISO had not received notice of (a) patent(s), which may be required to implement this document. However, implementers are cautioned that this may not represent the latest information, which may be obtained from the patent database available at <https://patents.iec.ch> and [www.iso.org/patents](http://www.iso.org/patents). IEC and ISO shall not be held responsible for identifying any or all such patent rights.

IEC 80601-2-71 has been prepared by a Joint Working Group of IEC subcommittee 62D: Particular medical equipment, software, and systems, of IEC technical committee 62: Medical equipment, software, and systems, and ISO subcommittee SC3: Respiratory devices and related equipment used for patient care, of ISO technical committee 121: Anaesthetic and respiratory equipment. It is an International Standard.

This second edition cancels and replaces the first edition published in 2015. This edition constitutes a technical revision.

This edition includes the following significant technical changes with respect to the previous edition:

- a) alignment with IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012, IEC 60601-1:2005/AMD2:2020, IEC 60601-1-8:2006, IEC 60601-1-8:2006/AMD1:2012, IEC 60601-1-8:2006/AMD2:2020, IEC 60601-1-2:2014, IEC 60601-1-2:2014/AMD1:2020, IEC 60601-1-6:2010, IEC 60601-1-6:2010/AMD1:2013 and IEC 60601-1-6:2010/AMD2:2020;
- b) added requirements for ESSENTIAL PERFORMANCE;
- c) added requirements for PRIMARY OPERATING FUNCTIONS;
- d) added requirements for protection against excessive temperatures;
- e) added requirements for the display legibility for OPERATORS wearing personal protective equipment;
- f) harmonization with ISO 20417, where appropriate.

This publication is published as a double logo standard.

The text of this International Standard is based on the following documents of IEC:

Draft	Report on voting
62D/2169/FDIS	62D/2196/RVD

Full information on the voting for its approval can be found in the report on voting indicated in the above table.

The language used for the development of this International Standard is English.

This document was drafted in accordance with ISO/IEC Directives, Part 2, and developed in accordance with ISO/IEC Directives, Part 1 and ISO/IEC Directives, IEC Supplement, available at [www.iec.ch/members\\_experts/refdocs](http://www.iec.ch/members_experts/refdocs). The main document types developed by IEC are described in greater detail at [www.iec.ch/publications](http://www.iec.ch/publications).

In this document, the following print types are used:

- requirements and definitions: roman type.
- *test specifications: italic type.*
- informative material appearing outside of tables, such as notes, examples and references: in smaller type. Normative text of tables is also in a smaller type.
- TERMS defined in Clause 3 of IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012 AND IEC 60601-1:2005/AMD2:2020, IN THIS DOCUMENT OR AS NOTED: SMALL CAPITALS.

In referring to the structure of this document, the term

- "clause" means one of the seventeen numbered divisions within the table of contents, inclusive of all subdivisions (e.g. Clause 7 includes subclauses 7.1, 7.2, etc.);
- "subclause" means a numbered subdivision of a clause (e.g. 7.1, 7.2 and 7.2.1 are all subclauses of Clause 7).

References to clauses within this document are preceded by the term "Clause" followed by the clause number. References to subclauses within this particular standard are by number only.

In this document, the conjunctive "or" is used as an "inclusive or" so a statement is true if any combination of the conditions is true.

The verbal forms used in this document conform to usage described in Clause 7 of the ISO/IEC Directives, Part 2. For the purposes of this document, the auxiliary verb:

- "shall" means that compliance with a requirement or a test is mandatory for compliance with this document;
- "should" means that compliance with a requirement or a test is recommended but is not mandatory for compliance with this document;
- "may" is used to describe a permissible way to achieve compliance with a requirement or test.

An asterisk (\*) as the first character of a title or at the beginning of a paragraph or table title indicates that there is guidance or rationale related to that item in Annex AA.

A list of all parts of the IEC 80601 series, published under the general title *Medical electrical equipment*, can be found on the IEC website.

The committee has decided that the contents of this document will remain unchanged until the stability date indicated on the IEC website under [webstore.iec.ch](https://webstore.iec.ch) in the data related to the specific document. At this date, the document will be

- reconfirmed,
- withdrawn, or
- revised.

NOTE The attention of the users of this document is drawn to the fact that equipment manufacturers and testing organizations may need a transitional period following publication of a new, amended or revised IEC or ISO publication in which to make products in accordance with the new requirements and to equip themselves for conducting new or revised tests. It is the recommendation of the committees that the content of this publication be adopted for implementation nationally not earlier than 3 years from the date of publication.

[SIST EN IEC 80601-2-71:2025](https://standards.iteh.ai/catalog/standards/sist/6e3ca6cb-172e-442e-8676-a698673c16ba/sist-en-iec-80601-2-71-2025)

<https://standards.iteh.ai/catalog/standards/sist/6e3ca6cb-172e-442e-8676-a698673c16ba/sist-en-iec-80601-2-71-2025>

## INTRODUCTION

The minimum safety requirements specified in this document are considered to provide for a practical degree of safety in the operation of FUNCTIONAL NIRS EQUIPMENT.

The requirements are followed by specifications for the relevant tests.

A "Particular guidance and rationale" text giving some explanatory notes, where appropriate, about the more important requirements is included in Annex AA. It is considered that knowledge of the reasons for these requirements will not only facilitate the proper application of the standard but will, in due course, expedite any revision necessitated by changes in clinical practice or as a result of developments in technology. However, this annex does not form part of the requirements of this document.

**iTeh Standards**  
**(<https://standards.iteh.ai>)**  
**Document Preview**

[SIST EN IEC 80601-2-71:2025](https://standards.iteh.ai/catalog/standards/sist/6e3ca6cb-172e-442e-8676-a698673c16ba/sist-en-iec-80601-2-71-2025)

<https://standards.iteh.ai/catalog/standards/sist/6e3ca6cb-172e-442e-8676-a698673c16ba/sist-en-iec-80601-2-71-2025>

## MEDICAL ELECTRICAL EQUIPMENT –

### Part 2-71: Particular requirements for the basic safety and essential performance of functional near-infrared spectroscopy (NIRS) equipment

#### 201.1 Scope, object and related standards

Clause 1 of IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012 and IEC 60601-1:2005/AMD2:2020 applies, except as follows:

##### 201.1.1 Scope

###### *Replacement:*

This part of IEC 80601 applies to the BASIC SAFETY and ESSENTIAL PERFORMANCE of FUNCTIONAL NIRS EQUIPMENT, as defined in 201.3.205, intended to be used by itself, or as a part of an ME SYSTEM hereinafter referred to as ME EQUIPMENT.

HAZARDS inherent in the intended physiological function of ME EQUIPMENT or ME SYSTEMS within the scope of this document are not covered by specific requirements in this document except in IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012 and IEC 60601-1:2005/AMD2:2020, 7.2.13 and 8.4.1.

NOTE Additional information can be found in IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012 and IEC 60601-1:2005/AMD2:2020, 4.2.

This document is not applicable to

- equipment for the measurement of oxygen saturation of the haemoglobin in the micro vessels (capillaries, arterioles and venules), i.e. tissue oximeters;
- frequency-domain and time-domain equipment for functional near-infrared spectroscopy;
- equipment for the measurement of changes in the concentration of chromophores other than oxy- and deoxy-haemoglobin;
- equipment for the measurement of changes in the concentration of oxy- and deoxy-haemoglobin in tissues other than the brain.

This document does not specify the requirements for:

- cerebral tissue oximeter equipment, which are given in ISO 80601-2-85 [1]<sup>1</sup>; and
- pulse oximeter equipment, which are given in ISO 80601-2-61 [2].

##### 201.1.2 Object

###### *Replacement:*

The object of this document is to establish particular BASIC SAFETY and ESSENTIAL PERFORMANCE requirements for FUNCTIONAL NIRS EQUIPMENT as defined in 201.3.205.

NOTE This document has been prepared to address the relevant essential principles [3] and labelling principles [4] of the International Medical Devices Regulators Forum (IMDRF) as indicated in Annex DD.

<sup>1</sup> Numbers in square brackets refer to the Bibliography.